

Date:

Report Type:

24.11.2024

Maintenance

Customer:

Name: Dr. Shaban	Location: Sul
------------------	---------------

Machine & Inspection:

Name	Type	Model	Serial Number	Lamp Shoots	✓/✗
Zimmer	Aircooling	CryoZet	ACSL2111327	-	X

Report:

The motor was stopped

Actions:

I change the coils

Price:

500 -> 300

Customer Name & signature

(Signature)

Engineer name & signature:

(Signature)